




FRAUNHOFER-INSTITUTE FOR INTERFACIAL ENGINEERING AND BIOTECHNOLOGY IGB

Proof of Conformity

In vitro Cytotoxicity according to DIN ISO 10993-5, 2009

Test Laboratory	Test Laboratory Cell and Tissue Engineering Fraunhofer Institute for Interfacial Engineering and Biotechnology IGB Nobelstrasse 12 70569 Stuttgart
Company	3AC POLYCASA GmbH  
Offer Number	11 16 014
Order Date	23.02.2016
Test method	<i>In vitro</i> Cytotoxicity according to DIN ISO 10993-5, 2009
Period of testing	02.03. to 08.04.2016
Test item	POLYCASA SAN, extruded styrene acrylic nitrile sheets 
Internal Code	ZS / 2 / 11 16 014 / 4
Extraction	conform with DIN ISO 10993-12 Extraction relationship 3 cm ² / ml 72h ± 2h at 37°C ± 1°C and 5% CO ₂

Result of extract testing

The extract of the test item was tested non cytotoxic.

18.04.2016 Dpl. Biol. (t.o.) S. Thiele

Date, Title and Signature of Testing Leader